

List Certificate

Dual temperature probes, part number 02 101581 00



VARIAN

Varian Australia Pty Ltd
679 Springvale Road
Mulgrave VIC 3170
Australia

Operator: Van D. T. T.	Date: 13.9.10
Serial number: UV1009M382	

Temperature accuracy

	Reference temperature		Probe temperature		Customer specification
	Target	Actual		°C	
Probe 1	0.0	-0.001 °C	-0.01	°C	± 0.3 °C
	100.0	86.38 °C	86.37	°C	± 0.3 °C
	50.0	49.09 °C	49.08	°C	± 0.3 °C
Probe 2	0.0	-0.001 °C	0.00	°C	± 0.3 °C
	100.0	86.38 °C	86.37	°C	± 0.3 °C
	50.0	49.09 °C	49.08	°C	± 0.3 °C